

Queen's University Environmental Health & Safety

Date Issued:
January 2006

Page No.:
1

Document No.:
SOP-Biosafety - 01

Revision:
1.0

Subject:
Vacuum & Aspiration Equipment



Date Issued: January 2006	Page No.: 2	Document No.: SOP-Biosafety - 01
Revision: 1.0	Subject: Vacuum & Aspiration Equipment	

Where possible, the secondary containment flask should be placed higher than the experiment so that liquid accidentally aspirated into the flask can drain back to the primary reservoir when the connection to the vacuum line is broken.